



INDIAN INSTITUTE OF TECHNOLOGY BOMBAY
MATERIALS MANAGEMENT DIVISION
Powai, Mumbai - 400076

21 January, 2026

Corrigendum – II

PR No. 1000051768

RFx No. 6100002561

Semiconductor Curve Tracer System

Sr. No.	Item Description	Previous Clause	Changed Clause
1.	Medium-Voltage (MV) unit	1.Voltage ≥ 200 V 2.Current ≥ 2 A pulse & ≥ 200 mA DC 3.Resolution ≤ 250 fA 4.Synchronised gate/drain pulsing	1.Voltage ≥ 200 V 2.Current ≥ 1 A pulse & ≥ 200 mA DC 3.Resolution ≤ 250 fA 4.Synchronised gate/drain pulsing
2.	Measurement Features	1.Static & pulsed IV 2.Transfer/output curves 3.Gate charge 4.Leakage current measurements 5.C-V measurements	1.Static & pulsed IV 2.Transfer/output curves 3.Leakage current measurements


Deputy Registrar

21/01/26

Materials Management Division

Need
21/1/2026